

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/01/2020

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:17
INTERNAL STANDARD	VERIFIED	23:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:18
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:19
BLANK TEST	0.00	23:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:20
BLANK TEST	0.00	23:21

