

SUPERVISOR TEST

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 12/29/2020

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:06
INTERNAL STANDARD	VERIFIED	19:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:06
BLANK TEST	0.00	19:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:08
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:09
BLANK TEST	0.00	19:10

