

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

ate: 01/05/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:52
INTERNAL STANDARD	VERIFIED	03:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:54
BLANK TEST	0.00	03:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:55
BLANK TEST	0.00	03:56

