

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 07/28/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:48
INTERNAL STANDARD	VERIFIED	16:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:48
BLANK TEST	0.00	16:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:49
BLANK TEST	0.00	16:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:51
BLANK TEST	0.00	16:52

