SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 04/03/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:04
INTERNAL STANDARD	VERIFIED	14:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:05
BLANK TEST	0.00	14:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:06
BLANK TEST	0.00	14:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:07
BLANK TEST	0.00	14:08