

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/17/2022

OPERATOR NAME:
THOMAS J ZOVISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:01
INTERNAL STANDARD	VERIFIED	02:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:01
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:03
BLANK TEST	0.00	02:04

