CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 18360, dated December 12, 2018

I, Jennifer F Limoges, Associate Director of Forensic Science, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivisions a and c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record, once created, is stored in an electronic format that cannot, thereafter, be altered or modified. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.

Digitally signed under January F. Limoges on 2018.12.13 13:44:03 -05'00'

Jennifer F Limoges

Associate Director of Forensic Science Forensic Investigation Center



ANDREW M. CUOMO Governor **GEORGE P. BEACH II** Superintendent

RAY A. WICKENHEISER Director

0.10% SIMULATOR SOLUTION RECORD

December 12, 2018 Report Date:

Manufacturer: Guth Laboratories, Inc.

Lot Number: 18360

Expiration Date: 5/28/2019

Analysis Results: 121 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, tested simulator solution lot number 18360 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

Digitally signed under ESRA by Carol S Miller on 2018.12.12 11:41:56 -05'00'

Carol S Miller

Forensic Scientist III

Breath Analysis Technical Supervisor/Operator #43419

Reviewed by:

Digitally signed under ESRA by Carrie A Kirkton Carri A. Kustra

2018.12.13 09:33:06 -05'00'

Carrie A Kirkton

Supervisor of Forensic Services

Breath Analysis Technical Supervisor