

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Washington County Sheriff's Dept
Manufacturer: INTOXIMETERS INC **Model:** ALCO-SENSOR FST
Serial Number: 245059
Date: July 07, 2020

Reference Solution Lot Number: 20040
Reference Solution Standard: 0.100 % at 34.00 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.098 %	33.99 degrees centigrade
2	0.098 %	34.00 degrees centigrade
3	0.097 %	33.99 degrees centigrade
4	0.098 %	34.00 degrees centigrade



Michael J. Canary
Highway Safety Equipment Technician
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Washington County Sheriff's Dept
Manufacturer: INTOXIMETERS INC
Model: ALCO-SENSOR FST
Serial Number: 245059
Date: July 07, 2020


1. General Appearance and condition of this device -----: OK

2. Check battery indicators -----: OK



3. Check zero flashing -----: OK

4. Record any repairs made to this instrument and list services provided: **VERIFIED CALIBRATION**



Michael J. Canary
Highway Safety Equipment Technician
Office of Public Safety

SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency WASHINGTON COUNTY SO		Today's Date 7-7-2020
Address 399 BROADWAY		City, State, ZIP FORT EDWARD NY 12828
Contact Person		Contact Telephone
Contact Email		Best Way to Reach You During Appointment Period

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle): NPAS DMT GUTH SIM Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPN A/S FST Other:		
Model Number FST	Instrument Serial Number 245059	
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -		Type of maintenance (check appropriate) <input type="checkbox"/> Calibration Check <input type="checkbox"/> Repair
Brief Description of Malfunction		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT SIM Tubes Keyboard Other:		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:
<i>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</i>		
Signature N/A NEW ISSUE		Date 7-7-2020

DCJS USE ONLY

INT INSP	LOT 20040
0.098 @ 33.99°C	0.098 @ 33.99°C
	0.098 @ 34.00°C
	0.097 @ 33.99°C
	0.098 @ 34.00°C